

LVD TEST & INSPECTION CERTIFICATE PAGE 1 OF 1

ISSUED BY:
CRANAGE EMC TESTING LIMITED
STABLE COURT
OAKLEY
MARKET DRAYTON
SHROPSHIRE, TF9 4AG

Certificate No: L7656C2 Date of Issue: 01 September 2014

This document certifies that the product (EUT) has been assessed for compliance with the stated standard as food thermometer, and specifically found to satisfy the applicable requirements of;

BS EN 13485: 2001

Thermometers for measuring the air and product temperature for the transport, storage and distribution of chilled, frozen, deep-frozen/quick-frozen food and ice cream – Tests, performance, suitability

This certificate covers the assessment of a representative test sample along with its technical documentation to the above standard only. Other standards have not been considered. Testing and construction details are retained in Cranage File L7656. When installed, the following considerations shall be made: -

- | | | | |
|--------------------------|----------------|--------------------------|----------------|
| 1. Application: | Chilled/frozen | 6. Accuracy Class: | 0.5 |
| 2. Usage type: | Internal temp. | 7. Response time: | 3 min |
| 3. Suitability: | Storage | 8. Operating Conditions: | -20°C to +30°C |
| 4. IP Rating: | IP65 | 9. Limiting Conditions: | -30°C to +50°C |
| 5. Climatic Environment: | E | | |

Tested & Inspected for: **Hanna Instruments Ltd,
Eden Way,
Pages Industrial Park
Leighton Buzzard
Beds. LU7 4AD**

Date of First Test: **31/07/14**
Date of Last Test: **26/08/14**

Manufacturer: **Hanna Instruments Ltd**
Model: **Checktemp 1**
Product Description: **Portable food thermometer**
Serial Number: **—**
Ratings: **As above**

Approved Signatories: M. Murray (Test Engineer)



APPROVED SIGNATORY

Verification Signatories: M. Luke (Senior Safety Engineer)



VERIFICATION SIGNATORY